

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: LH050042G

Date Tested: 19-Jun-17

Instrument Number: 12756

Mac Address: 00:0A:8D:00:31:D4

IP Address: 196.168.111.222

| Primary Standard | | | | | | |
|------------------|----------------|------------------|----------------|------------------|----------------|------------------|
| | 382 (F) | | 882 (F) | | 1382 (F) | |
| | Signal Voltage | Recorder display | Signal Voltage | Recorder display | Signal Voltage | Recorder display |
| Channel 1 | 2.784 | 382 | 6.011 | 883 | 9.239 | 1384 |
| Channel 2 | 2.781 | 382 | 6.013 | 883 | 9.247 | 1386 |
| Channel 3 | 2.788 | 382 | 6.016 | 882 | 9.247 | 1386 |
| Channel 4 | 2.792 | 384 | 6.025 | 884 | 9.243 | 1385 |
| Channel 5 | 2.798 | 383 | 6.024 | 883 | 9.248 | 1385 |
| Channel 6 | 2.794 | 382 | 6.012 | 884 | 9.244 | 134 |

Signal voltage directly from Fluke 714

| | | | | | | |
|------------|--------|-----|--------|-----|--------|------|
| Channel 7 | direct | 382 | direct | 883 | direct | 1383 |
| Channel 8 | direct | 381 | direct | 881 | direct | 1382 |
| Channel 9 | direct | 381 | direct | 882 | direct | 1382 |
| Channel 10 | direct | 382 | direct | 881 | direct | 1383 |
| Channel 11 | direct | 382 | direct | 882 | direct | 1383 |
| Channel 12 | direct | 381 | direct | 881 | direct | 1382 |

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:


Fluke 714 Thermocouple Calibrator SN: 2097001 Calibration Date: 13-Mar-17

Precision digital multimeter SN: 2142012 Calibration Date: 13-Mar-17

ProHeat 35 power supply used for signal output SN: ME040117 Calibration date: 19-Jun-17

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Technician: 

Signature

Recalibration due date: 19-Jun-18