

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: LJ030208G
 Instrument Number: 50942
 IP Address: 192.168.111.222

Date Tested: Dec 9/13
 Mac Address: 00:0A:8D:C6:FE

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.786	382	6.013	883	9.238	1383
Channel 2	2.786	382	6.013	882	9.239	1382
Channel 3	2.787	382	6.013	882	9.239	1382
Channel 4	2.787	382	6.013	882	9.239	1382
Channel 5	2.787	382	6.013	882	9.239	1382
Channel 6	2.787	382	6.013	882	9.239	1382

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: ^{AUG} Dec 9/13

Precision digital multimeter SN: 2142011

Calibration Date: ^{AUG} Dec 9/13

ProHeat 35 power supply used for signal output SN: MC430028


Calibration date: Dec 9/13

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Calibration due date: Dec 9/14

Technician: _____


 Signature