

Calibration Certificate

Unit Serial Number: LJ160024G

Date Tested: 8/21/2012

Instrument Number: 53936

Mac Address: 00:0A:8D:00:D2:B0

IP Address: 192.168.111.222

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.779	381	6.010	882	9.240	1382
Channel 2	2.779	381	6.010	882	9.233	1382
Channel 3	2.779	381	6.010	882	9.234	1381
Channel 4	2.779	381	6.010	882	9.236	1382
Channel 5	2.779	381	6.010	882	9.238	1382
Channel 6	2.779	381	6.010	882	9.240	1383

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: N/A

Calibration Date: N/A

Precision digital multimeter SN: 2036021

Calibration Date: 5/2/2012

ProHeat 35 power supply used for signal output SN: N/A

Calibration date: N/A

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Technician: Scott T

Calibration due date: 8/21/2013

Signature