

## CERTIFICATE OF CALIBRATION

Unit Model #:

195374

Unit Serial #:

LK220011G

Recorder Serial #:

US26605-1-9-2809-PL1

Instrument #:

64594

MAC Address:

00:0A:8D:00:FC:52

Sure Controls does hereby certify the above instrument was calibrated against standards maintained by Sure Controls and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

## Displayed Value

Recorder	Primary Standard Signal 382.0 ° F	Primary Standard Signal 882.0° F	Primary Standard Signal 1382.0° F
CHANNEL 1	382	882	1382
CHANNEL 2	382	882	1382
CHANNEL 3	382	882	1382
CHANNEL 4	382	882	1382
CHANNEL 5	382	882	1382
CHANNEL 6	382	882	1382

Calibration Date:

1/13/2014

Recalibration Date:

1/13/2015

Certified by:

Instrumentation Used:

Beamex Model No. MC2 Multi-Function Calibrator

Beamex Serial No. 13533

Sure Controls ID No. STD-39