

## Calibration Certificate

Unit Serial Number: MA280262G

Date Tested: 10/3/2017

Instrument Number: 96723

Mac Address: 00:0A:8D:01:79:D3

IP Address: 192.168.111.222

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.787	382	6.013	882	9.239	1382
Channel 2	2.787	382	6.013	882	9.239	1382
Channel 3	2.787	382	6.013	882	9.239	1382
Channel 4	2.787	382	6.013	882	9.239	1382
Channel 5	2.787	382	6.013	882	9.239	1382
Channel 6	2.787	382	6.013	882	9.239	1382

**Allowable difference in temp reading is +or- 5 Deg. F.**

### Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 1882015

Calibration Date: 11/8/2016

Precision digital multimeter SN: 2036021

Calibration Date: 11/8/2016

ProHeat 35 power supply used for signal output SN: N/A

Calibration date: N/A

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Technician: Scott T.

Calibration due date: 10/3/2018

Signature