

Red-D-Arc Welderrentals

Calibration Certificate

Miller ProHeat 35

Unit Serial Number: MC430028

Date Tested: Dec 9/13

Certified Calibration

Primary Standard						
	382° (F) 2.787VDC@RC9		882° (F) 6.013VDC@RC9		1382° (F) 9.239VDC@RC9	
	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display
TC1	2.794	383	6.021	883	9.241	1382
TC2	2.797	383	6.016	883	9.239	1382
TC3	2.802	383	6.027	884	9.244	1382
TC4	2.799	384	6.021	884	9.240	1383
TC5	2.798	384	6.025	884	9.247	1383
TC6	2.797	384	6.027	884	9.252	1383
Allowable	2.749-2.825	+/- 3° F	5.975-6.051	+/- 3° F	9.201-9.277	+/- 3° F

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: ^{Aug} ~~Dec~~ 9/13

Precision Digital Voltage Meter SN: 2142011

Calibration Date: ^{Aug} ~~Dec~~ 9/13


Machine re-calibration due date: Dec 9/14

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician: _____



Date: Dec 9/13

Signature