

# Red-D-Arc Welderrentals

## Calibration Certificate

### Miller ProHeat 35

Unit Serial Number: MD130155G

Date Tested: 07/10/15

#### Certified Calibration

Primary Standard						
	382° (F) 2.787VDC@RC9		882° (F) 6.013VDC@RC9		1382° (F) 9.239VDC@RC9	
	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display
TC1	2.789	383	6.017	882	9.242	1382
TC2	2.799	384	6.019	883	9.246	1383
TC3	2.802	384	6.022	883	9.252	1383
TC4	2.799	384	6.020	883	9.250	1383
TC5	2.794	383	6.025	883	9.249	1383
TC6	2.795	384	6.026	884	9.252	1383
Allowable	2.749-2.825	+/- 3° F	5.975-6.051	+/- 3° F	9.201-9.277	+/- 3° F

#### Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: 24/07/14

Precision Digital Voltage Meter SN: 2142011

Calibration Date: 24/07/14

Machine re-calibration due date: 07/10/16

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician: Guo Boden Date: 07-10-15

Signature